

# DORO Aluminium Head Rest System

DORO Aluminium Head Rest System for supine, prone, lateral and sitting position (for sitting position Crossbar Adaptor is required). Excellent stability and flexibility. 3-pin rigid cranial fixation. Cannot be used for any MRI Imaging. Fixing Adaptor 60146 required for Promerix and Practico

## Includes the items below:

101 001 001     *Doro QR3 Skull Clamp*  
100 300 200     *Doro Swivel Adaptor Aluminium*  
100 300 100     *Doro Base Unit Aluminium*



## Technical Specifications

Product Code	101001100

## Ota yhteyttä

**Merivaara Corp.**  
Tarmontie 2-4  
15860 HOLLOLA  
SUOMI

+358 3 3394 611

[merivaara@merivaara.com](mailto:merivaara@merivaara.com)

